

## Frequently Used Bruker Commands

Command	Description
ej	Turn on lifting Lock the spectrometer to the deuterium signal of your solvent
topshim	Shim the magnetic field to improve its homogeneity for better resolution
new	Create a new experiment with options by using current a new sequence. New sequence.224 O TD (a)